

The following FOI questions relate to the Trusts Sterile Services and Endoscopy Reprocessing facilities. Thank you very much for taking time to complete this request.

1. Where are the Trust's Sterile Services Department/s (SSD) located? If the service is located off-site, please provide the address. **Blackpool Teaching Hospital NHS Foundation Trust, Whinney Heys Road, Blackpool, FY3 8NR.**
2. Does the processing of flexible endoscopes take place in dedicated 'flexible endoscope decontamination unit/facility'? If not, how many clinical locations undertake these tasks? **YES**
3. Are both sterile service and the endoscopy reprocessing facilities accredited to MDD 93/42/EEC and ISO13485 and, do they meet the requirements of HBN13, 2004? Is the endoscopy reprocessing HTM0106 and JAG compliant?

Accreditation	Sterile Services		Endoscopy Reprocessing	
	Yes	No	Yes	No
MDD 93/42/EEC	✓			✓
ISO13485	✓		✓	
HBN13, 2004	✓			✓
HTM0106			✓	
JAG			✓	

4. Which Trust department do the Sterile Service Department and Endoscopy reprocessing facilities sit under, e.g. Facilities & Estates Department, Surgical Directorate, a Trust sub-co. Please will the Trust provide the current named budget holder and contact details.

	Sterile Services	Endoscopy
Department each sit under	Facilities	Unscheduled Care
Budget Holder Name	Dorothy Wardrope	Lynne Butler
Email contact	Dorothy.wardrope@nhs.net	Lynne.butler3@nhs.net

5. What equipment is currently installed in the Trust's Sterile Services Department to include washer disinfectors, steam sterilisers (autoclaves), and low temperature steriliser e.g. Sterrad, V-PRO etc.? Please complete the tables below:

**Washer Disinfectors**

Manufacturer e.g. Getinge, Belimed etc.	Model	Number of Machines	Number of Chambers	Install Date
Dekomed	Deko 2000	6	1	April 2005

**Steam Sterilisers (Autoclaves)**

Manufacturer e.g. Getinge, Belimed etc.	Model	Number of machines	Install Date
BMM Weston	Porous Load Series 3	4	2014 x1, 2015 x1, 2018 x2

**Low Temperature Sterilisers**

Manufacturer e.g. Sterrad	Model	Number of machines	Install Date
N/a			

6. For the equipment stated in Q5 above, please state the number of cycles completed over the last 12 months for all:

- a. Washer Disinfectors Annual Cycle Count = **10,410**
- b. Steam Sterilisers Annual Cycle Count = **12,615**
- c. Low-Temperature Sterilisers Annual Cycle Count = **N/A**

7. In the table below, please provide details (manufacturer, product name, and price) on the consumables e.g. washer disinfectant detergent, utilised in the equipment listed in Q8.

Equipment Type	Consumable	Manufacturer	Product Name & Code
Washer Disinfectant	Detergent #1	Serchem	Phoenix 1091
Washer Disinfectant	Detergent #2	Dekomed	Dekomed DET 001
Autoclave	Bowie Dick Test	Brownes Test Pack supplied by Steris	Brownes Test Pack EUF2212
Low-Temp Steriliser	Hydrogen Peroxide	N/a	N/a

8. In the tables below, please provide information detailing the location, manufacturer, model, quantity, and age, relating to the following equipment used in the decontamination of flexible endoscopes:

- a. Automatic Endoscope Reprocessors;
- b. Low-Temp Sterilisers e.g. Sterrad;
- c. Flexible Endoscope Drying Cabinets; and,

d. Reverse Osmosis Water System.

**a) Automatic Endoscope Reprocessors**

Location e.g. Endoscopy Department, Theatres, CSSD	Manufacturer	Model	Quantity	Age(s)
EDU	Olympus	ETD-Double	4	3
EDU	Olympus	ETD-Double	2	New

**b) Low-Temp Sterilisers**

Location e.g. Endoscopy Department, Theatres, CSSD	Manufacturer	Model	Quantity	Age(s)

**c) Drying Cabinets**

Location e.g. Endoscopy Department, Theatres, CSSD	Manufacturer	Model	Quantity	Age(s)
EDU	Olympus	EDC Plus	2	3

**d) RO Water System**

Manufacturer	Model	Age
EDU	EWS/EndoTherm Duo	3

9. For the equipment stated in Q8 above, please state the number of cycles completed over the last 12 months:

a. Automatic Endoscope Reprocessors Annual Cycle Count = **26,762**

**b. Low-Temperature Sterilisers Annual Cycle Count = 0**

10. In the table below, please provide details (manufacturer and product name/code) on the consumables e.g. Automatic Endoscope Reprocessor detergent, utilised in the equipment listed in Q11.

Equipment Type	Consumable	Manufacturer	Product Name/Code
Automatic Endoscope Reprocessor	Detergent #1	Ecolab	EndoDis EndoAct EndoDet
Automatic Endoscope Reprocessor	Detergent #2		
Low-Temp Steriliser	Hydrogen Peroxide		

11. Have you planned investment in either the Sterile Services or Endoscopy reprocessing facility - e.g. Autoclaves, Automated washer disinfectors machines, reverse osmosis plant, boilers etc.? If so, please list the required investment and has provision been made within the FY20 and FY21 budgets?

Proposed Investment	Sterile Services	Endoscopy Reprocessing
FY20	N/A	N/A
FY21	Unsure of investment – 6 x WD's to be discussed for FY21	N/A